


## *Vacuum protection in the Biosafety Cabinet*

- Place hydrophobic filter between overflow flask and pump
  - Whatman Vacu-guard in-line disk filters
  - Pall Gelman Vacushield
- Turn off vacuum when not in use



## **Vacuum Protection inside the BSC prevents overflow from entering the main vacuum line**



Place hydrophobic filter between overflow flask and pump  
 Whatman Vacu-guard in-line disk filters  
 Pall Gelman Vacushield  
 Turn off vacuum when not in use